



**Relative Accuracy Test Audit
Test Report**

**Michigan State University
T.B. Simon Power Plant
Combustion Turbine 6
East Lansing, Michigan
February 14, 2017**

**Report Submittal Date
April 26, 2017**

© Copyright 2017
All rights reserved in
Mostardi Platt

Project No. M170605E

RECEIVED

MAY 02 2017

AIR QUALITY DIV.

1.0 EXECUTIVE SUMMARY

MOSTARDI PLATT conducted a Continuous Emissions Monitoring System (CEMS) Relative Accuracy Test Audit (RATA) test program for Michigan State University at the T.B. Simon Power Plant in East Lansing, Michigan, on the Combustion Turbine 6 on February 14, 2017. This report summarizes the results of the test program and test methods.

The test location, test date, and test parameters are summarized below.

TEST INFORMATION		
Test Location	Test Date	Test Parameters
Combustion Turbine 6	February 14, 2017	Carbon Monoxide (CO), Oxygen (O ₂), and Nitrogen Oxides (NO _x)

The purpose of the test program was to demonstrate the relative accuracies of the Combustion Turbine 6 CO, O₂, and NO_x analyzers during the specified operating condition. The test results from this test program indicate that each CEMS meets the United States Environmental Protection Agency (USEPA) annual performance specification for relative accuracy as published in 40 Code of Federal Regulations Part 60 (40CFR60).

RATA RESULTS					
Test Location	Date	Parameter	Units	Relative Accuracy Acceptance Criteria	Relative Accuracy (RA)
Combustion Turbine 6	2/14/17	NO _x	lb/mmBtu	≤ 20.0 % of the mean reference value	2.26%
		NO _x	ppmvd	≤ 20.0 % of the mean reference value	4.19%
		O ₂	% dry	≤ 20.0 % of the mean reference value	1.07%
		CO	ppmvd	≤ 10.0 % of the mean reference value	2.42%
		CO	lb/mmBtu	≤ 10.0 % of the mean reference value	0.70%

The gas cylinders used to perform the RATA are summarized below.

GAS CYLINDER INFORMATION				
Parameter	Gas Vendor	Cylinder Serial Number	Cylinder Value	Expiration Date
NO _x	Airgas	CC343413	0.0 ppm	7/11/2024
NO _x	Airgas	CC216539	47.06 ppm	11/7/2019
NO _x	Airgas	CC490411	92.92 ppm	10/31/2024
O ₂	Airgas	CC216539	0.0 %	11/7/2019
O ₂	Airgas	CC347269	9.082 %	6/13/2024
O ₂	Airgas	CC416859	22.19 %	5/14/2023
CO	Airgas	CC347269	0.0 ppm	6/13/2024
CO	Airgas	CC343413	50.31 ppm	7/11/2024
CO	Airgas	CC486880	88.98 ppm	7/7/2024

The identifications of the individuals associated with the test program are summarized below.

TEST PERSONNEL INFORMATION		
Location	Address	Contact
Test Facility	Michigan State University 354 Service Rd East Lansing, MI 48824	Mr. Rick Johnson Electrical Engineer (517) 884-7108 (phone) rjohnson@ipf.msu.edu
Testing Company Supervisor	Mostardi Platt 888 Industrial Drive Elmhurst, Illinois 60126	Mr. Stuart L. Burton Senior Project Manager 630-993-2100 (phone) sburton@mp-mail.com
Testing Company Personnel		Mr. David Dixon Test Technician

2.0 TEST METHODOLOGY

Emission testing was conducted following the United States Environmental Protection Agency (USEPA) methods specified in 40CFR60, Appendix A in addition the Mostardi Platt Quality Manual. Schematics of the test section diagram and sampling train used are included in Appendix A and B respectively. Calculation nomenclature are included in Appendix C. Copies of analyzer print-outs for each test run are included in Appendix D. CEM data and process data as provided by Michigan State University are included in Appendix E.

The following methodologies were used during the test program:

Method 3A Oxygen (O₂) Determination

Stack gas O₂ concentrations were determined in accordance with USEPA Method 3A. A Servomex analyzer was used to determine the O₂ concentrations in the manner specified in the

Method. The instrument has a paramagnetic detector and the O₂ operates in the nominal range of 0% to 25% with the specific range determined by the high-level calibration gas of 22.19%. High-range calibrations were performed using USEPA Protocol gas. Zero nitrogen (a low ppm pollutant in balance nitrogen calibration gases) was introduced during other instrument calibrations to check instrument zero. High- and a mid-range % O₂ levels in balance nitrogen were also introduced. Zero and mid-range calibrations were performed using USEPA Protocol gas after each test run. Copies of the gas cylinder certifications are found in Appendix G. This testing met the performance specifications as outlined in the Method.

Method 7E Nitrogen Oxides (NO_x) Determination

Stack gas NO_x concentrations and emission rates were determined in accordance with USEPA Method 7E, 40CFR60, Appendix A. A Thermo Scientific Model 42i-D Chemiluminescence Nitrogen Oxides Analyzer was used to determine nitrogen oxides concentrations, in the manner specified in the Method. The instrument operated in the nominal range of 0 ppm to 100 ppm with the specific range determined by the high-level span calibration gas of 92.92 ppm.

The Model 42i-D High Level is based on the principle that nitric oxide (NO) and ozone (O₃) react to produce a characteristic luminescence with an, intensity linearly proportional to the NO concentration. Infrared light emission results when electronically excited nitrogen dioxide (NO₂) molecules decay to lower energy states. Specifically,



NO₂ must first be transformed into NO before it can be measured using the chemiluminescent reaction. NO₂ is converted to NO by a molybdenum NO₂-to-NO converter heated to about 612.6 C. The ambient air sample is drawn into the Model 42i-D High Level through the sample bulkhead. The sample flows through a particulate filter, a capillary, and then to the mode solenoid valve. The solenoid valve routes the sample either straight to the reaction chamber (NO mode) or through the NO₂-to-NO converter and then to the reaction chamber (NO_x mode).

Dry air enters the Model 42i-D High Level through the dry air bulkhead, through a flow sensor, and then through a silent discharge ozonator. The ozonator generates the necessary ozone concentration needed for the chemiluminescent reaction. The ozone reacts with the NO in the ambient air sample to produce electronically excited NO₂ molecules. A photomultiplier tube (PMT) housed in a thermoelectric cooler detects the NO₂ luminescence.

The NO and NO_x concentrations calculated in the NO and NO_x modes are stored in memory. The difference between the concentrations is used to calculate the NO₂ concentration. The Model 42i-D High Level outputs NO, NO₂, and NO_x concentrations to both the front panel display and the analog outputs.

Stack gas was delivered to the analyzer via a Teflon® sampling line, heated to a minimum temperature of 250°F. Excess moisture in the stack gas was removed using a refrigerated condenser. The entire system was calibrated in accordance with the Method, using certified calibration gases introduced at the probe, before and after each test run. This testing met the performance specifications as outlined in the Method.

A list of calibration gases used and the results of all calibration and other required quality assurance checks are found in Appendix F. Copies of the gas cylinder certifications are found in Appendix G. The NO₂ to NO converter test can be found in Appendix H. This testing met the performance specifications as outlined in the Method.

Method 10 Carbon Monoxide (CO) Determination

Stack gas CO concentrations and emission rates were determined in accordance with USEPA Method 16, 40CFR60, Appendix A. A Thermo Scientific Model 48i Gas Filter Correlation Analyzer was used to determine carbon monoxide concentrations, in the manner specified in the Method. The instrument operated in the nominal range of 0 ppm to 100 ppm with the specific range determined by the high-level span calibration gas of 88.98 ppm.

The Model 48i operates on the principle that CO absorbs infrared radiation at a wavelength of 4.6 microns. Because infrared absorption is a non-linear measurement technique, it is necessary to transform the basic analyzer signal into a linear output. The Model 48i uses an internally stored calibration curve to accurately linearize the instrument output over any range up to a concentration of 16,000 ppm. The sample is drawn into the Model 48i through the sample bulkhead. The sample flows through the optical bench. Radiation from an infrared source is chopped and then passed through a gas filter alternating between CO and N₂. The radiation then passes through a narrow bandpass interference filter and enters the optical bench where absorption by the sample gas occurs. The infrared radiation then exits the optical bench and falls on an infrared detector. The CO gas filter acts to produce a reference beam which cannot be further attenuated by CO in the sample cell. The N₂ side of the filter wheel is transparent to the infrared radiation and therefore produces a measurement beam which can be absorbed by CO in the cell. The chopped detector signal is modulated by the alternation between the two gas filters with an amplitude related to the concentration of CO in the sample cell. Other gases do not cause modulation of the detector signal since they absorb the reference and measure beams equally. Thus, the GFC system responds specifically to CO. The Model 48i outputs the CO concentration to the front panel display, the analog outputs, and also makes the data available over the serial or Ethernet connection.

Stack gas was delivered to the analyzer via a Teflon® sampling line, heated to a minimum temperature of 250°F. Excess moisture in the stack gas was removed using a refrigerated condenser. The entire system was calibrated in accordance with the Method, using certified calibration gases introduced at the probe, before and after each test run.

A list of calibration gases used and the results of all calibration and other required quality assurance checks are found in Appendix F. Copies of the gas cylinder certifications are found in Appendix G. This testing met the performance specifications as outlined in the Method.

3.0 TEST RESULT SUMMARIES

Client: Michigan State University						Location: Unit 6 Stack Low Load			
Facility: T.B. Simon Power Plant						Date: 2/14/17			
Project #: M170605						Test Method: 7E, 3A			
Fuel Type: Natural Gas						Fuel Factor: 8710			
O2 based NOx lb/mmBtu RATA									
CEM Monitor Information									
NO _x Monitor/Model:				API 200EM		NO _x Serial # :		76421-383 (139)	
O2 Monitor/Model:				Servomex 1440D		O2 Serial # :		767078-381 (198)	
1=accept 0=reject	Test Run	KPPH	Test Date	Start Time	End Time	RM NO _x lb/MMBtu	CEM NO _x lb/MMBtu	(RM-CEM) Difference (di)	(RM-CEM) Difference ² (di ²)
1	1	50	02/14/17	08:50	09:10	0.038	0.038	0.000	0.000
1	2	50	02/14/17	09:30	09:50	0.037	0.037	0.000	0.000
1	3	50	02/14/17	10:12	10:32	0.038	0.038	0.000	0.000
1	4	50	02/14/17	11:00	11:20	0.037	0.037	0.000	0.000
1	5	50	02/14/17	11:46	12:06	0.037	0.037	0.000	0.000
1	6	50	02/14/17	12:34	12:54	0.038	0.037	0.001	0.000
1	7	50	02/14/17	13:30	13:50	0.038	0.037	0.001	0.000
1	8	50	02/14/17	14:08	14:28	0.038	0.037	0.001	0.000
0	9	50	02/14/17	14:50	15:10	0.039	0.037	0.002	0.000
1	10	50	02/14/17	15:32	15:52	0.038	0.037	0.001	0.000
0	11	50	02/14/17	16:10	16:30	0.040	0.037	0.003	0.000
0	12	50	02/14/17	16:50	17:10	0.039	0.037	0.002	0.000
n						9			
t(0.975)						2.306			
Mean Reference Method Value						0.038		RM avg	
Mean CEM Value						0.037		CEM avg	
Sum of Differences						0.004		di	
Mean Difference						0.000		d	
Sum of Differences Squared						0.000		di ²	
Standard Deviation						0.001		sd	
Confidence Coefficient 2.5% Error (1-tail)						0.000		cc	
Relative Accuracy						2.26		RA	

Client: Michigan State University						Location: Unit 6 Stack Low Load			
Facility: T.B. Simon Power Plant						Date: 2/14/17			
Project #: M170605						Test Method: 7E			
NO _x ppmvd RATA									
CEM Monitor Information									
NO _x Monitor/Model:				API 200EM		NO _x Serial # :		76421-383 (139)	
1=accept 0=reject	Test Run	KPPH	Test Date	Start Time	End Time	RM NO _x ppmvd	CEM NO _x ppmvd	(RM-CEM) Difference (di)	(RM-CEM) Difference ² (di ²)
1	1	50	02/14/17	08:50	09:10	10.9	10.7	0.2	0.0
1	2	50	02/14/17	09:30	09:50	10.8	10.6	0.2	0.0
1	3	50	02/14/17	10:12	10:32	10.9	10.8	0.1	0.0
1	4	50	02/14/17	11:00	11:20	10.7	10.5	0.2	0.0
1	5	50	02/14/17	11:46	12:06	10.6	10.3	0.3	0.1
1	6	50	02/14/17	12:34	12:54	10.8	10.5	0.3	0.1
1	7	50	02/14/17	13:30	13:50	10.9	10.3	0.6	0.4
1	8	50	02/14/17	14:08	14:28	11.0	10.5	0.5	0.3
0	9	50	02/14/17	14:50	15:10	11.4	10.5	0.9	0.8
1	10	50	02/14/17	15:32	15:52	10.9	10.4	0.5	0.3
0	11	50	02/14/17	16:10	16:30	11.3	10.3	1.0	1.0
0	12	50	02/14/17	16:50	17:10	11.3	10.3	1.0	1.0
n						9			
t(0.975)						2.306			
Mean Reference Method Value						10.833		RM avg	
Mean CEM Value						10.511		CEM avg	
Sum of Differences						2.900		di	
Mean Difference						0.322		d	
Sum of Differences Squared						1.170		di ²	
Standard Deviation						0.172		sd	
Confidence Coefficient 2.5% Error (1-tail)						0.132		cc	
Relative Accuracy						4.19		RA	

Client: Michigan State University						Location: Unit 6 Stack Low Load			
Facility: T.B. Simon Power Plant						Date: 2/14/17			
Project #: M170605						Test Method: 3A			
O ₂ % (dry) RATA									
CEM Monitor Information									
O ₂ Monitor/Model:				Servomex 1440D		O ₂ Serial # :		767078-381 (198)	
1=accept 0=reject	Test Run	KPPH	Test Date	Start Time	End Time	RM O ₂ % (dry)	CEM O ₂ % (dry)	(RM-CEM) Difference (di)	(RM-CEM) Difference ² (di ²)
1	1	50	02/14/17	08:50	09:10	14.6	14.7	-0.1	0.01
1	2	50	02/14/17	09:30	09:50	14.6	14.7	-0.1	0.01
1	3	50	02/14/17	10:12	10:32	14.6	14.7	-0.1	0.01
1	4	50	02/14/17	11:00	11:20	14.6	14.7	-0.1	0.01
1	5	50	02/14/17	11:46	12:06	14.7	14.8	-0.1	0.01
1	6	50	02/14/17	12:34	12:54	14.7	14.7	0.0	0.00
1	7	50	02/14/17	13:30	13:50	14.6	14.8	-0.2	0.04
1	8	50	02/14/17	14:08	14:28	14.6	14.8	-0.2	0.04
0	9	50	02/14/17	14:50	15:10	14.6	14.8	-0.2	0.04
0	10	50	02/14/17	15:32	15:52	14.6	14.8	-0.2	0.04
1	11	50	02/14/17	16:10	16:30	14.7	14.8	-0.1	0.01
0	12	50	02/14/17	16:50	17:10	14.6	14.8	-0.2	0.04
n						9			
t(0.975)						2.306			
Mean Reference Method Value						14.633		RM avg	
Mean CEM Value						14.744		CEM avg	
Sum of Differences						-1.000		di	
Mean Difference						-0.111		d	
Sum of Differences Squared						0.140		di ²	
Standard Deviation						0.060		sd	
Confidence Coefficient 2.5% Error (1-tail)						0.046		cc	
Relative Accuracy						1.07		RA	

Client: Michigan State University
 Facility: T.B. Simon Power Plant
 Project #: M170605

Location: Unit 6 Stack Low Load

Date: 2/14/17

Test Method: 10

CO ppmvd RATA
CEM Monitor Information

CO Monitor/Model:				Teledyne 300E		CO Serial # :		75478-380 (178)	
1=accept 0=reject	Test Run	KPPH	Test Date	Start Time	End Time	RM CO ppmvd	CEM CO ppmvd	(RM-CEM) Difference (di)	(RM-CEM) Difference ² (di ²)
1	1	50	02/14/17	08:50	09:10	32.4	31.7	0.7	0.49
1	2	50	02/14/17	09:30	09:50	33.6	32.8	0.8	0.64
1	3	50	02/14/17	10:12	10:32	32.9	32.1	0.8	0.64
1	4	50	02/14/17	11:00	11:20	33.0	32.5	0.5	0.25
1	5	50	02/14/17	11:46	12:06	33.1	32.4	0.7	0.49
1	6	50	02/14/17	12:34	12:54	32.7	31.9	0.8	0.64
1	7	50	02/14/17	13:30	13:50	32.2	31.6	0.6	0.36
1	8	50	02/14/17	14:08	14:28	32.2	31.5	0.7	0.49
1	9	50	02/14/17	14:50	15:10	32.0	31.2	0.8	0.64
0	10	50	02/14/17	15:32	15:52	32.2	31.2	1.0	1.00
0	11	50	02/14/17	16:10	16:30	31.9	30.8	1.1	1.21
0	12	50	02/14/17	16:50	17:10	32.3	31.2	1.1	1.21
n						9			
t(0.975)						2.306			
Mean Reference Method Value						32.678	RM avg		
Mean CEM Value						31.967	CEM avg		
Sum of Differences						6.400	di		
Mean Difference						0.711	d		
Sum of Differences Squared						4.640	di ²		
Standard Deviation						0.105	sd		
Confidence Coefficient 2.5% Error (1-tail)						0.081	cc		
Relative Accuracy						2.42	RA		

Client: Michigan State University
Facility: T.B. Simon Power Plant
Project #: M170605
Fuel Type: Natural Gas

Location: Unit 6 Stack Low Load
Date: 2/14/17
Test Method: 10, 3A
Fuel Factor: 8710

O2 based CO lb/mmBtu RATA

CEM Monitor Information

CO Monitor/Model:		Teledyne 300E		CO Serial # :		75478-380 (178)			
O2 Monitor/Model:		Servomex 1440D		O2 Serial # :		767078-381 (198)			
1=accept 0=reject	Test Run	KPPH	Test Date	Start Time	End Time	RM CO lb/MMBtu	CEM CO lb/MMBtu	(RM-CEM) Difference (di)	(RM-CEM) Difference ² (di ²)
1	1	50	02/14/17	08:50	09:10	0.068	0.068	0.000	0.000
0	2	50	02/14/17	09:30	09:50	0.071	0.070	0.001	0.000
1	3	50	02/14/17	10:12	10:32	0.069	0.069	0.000	0.000
1	4	50	02/14/17	11:00	11:20	0.069	0.069	0.000	0.000
1	5	50	02/14/17	11:46	12:06	0.071	0.070	0.001	0.000
1	6	50	02/14/17	12:34	12:54	0.070	0.069	0.001	0.000
1	7	50	02/14/17	13:30	13:50	0.068	0.069	-0.001	0.000
1	8	50	02/14/17	14:08	14:28	0.068	0.068	0.000	0.000
1	9	50	02/14/17	14:50	15:10	0.067	0.068	-0.001	0.000
1	10	50	02/14/17	15:32	15:52	0.068	0.068	0.000	0.000
0	11	50	02/14/17	16:10	16:30	0.068	0.067	0.001	0.000
1	12	50	02/14/17	16:50	17:10	0.068	0.068	0.000	0.000
n						10			
t(0.975)						2.262			
Mean Reference Method Value						0.069	RM avg		
Mean CEM Value						0.069	CEM avg		
Sum of Differences						0.000	di		
Mean Difference						0.000	d		
Sum of Differences Squared						0.000	di²		
Standard Deviation						0.001	sd		
Confidence Coefficient 2.5% Error (1-tail)						0.000	cc		
Relative Accuracy						0.70	RA		

4.0 CERTIFICATION

MOSTARDI PLATT is pleased to have been of service to Michigan State University. If you have any questions regarding this test report, please do not hesitate to contact us at 630-993-2100.

CERTIFICATION

As the program manager, I hereby certify that this test report represents a true and accurate summary of emissions test results and the methodologies employed to obtain those results. The test program was performed in accordance with the test methods and the Mostardi Platt Quality Manual, as applicable.

MOSTARDI PLATT



Stuart L. Burton

Program Manager



Jeffrey M. Crivare

Quality Assurance