



STATE OF MICHIGAN
 DEPARTMENT OF ENVIRONMENTAL QUALITY
 LANSING



GRETCHEN WHITMER
 GOVERNOR

LIESL EICHLER CLARK
 DIRECTOR

January 4, 2019

Mr. Bryan Curry
 Viant Medical, Inc.
 520 Watson SW
 Grand Rapids, Michigan 49504

SRN: N0795, Kent County

Dear Mr. Curry:

VIOLATION NOTICE

On January 4, 2019, the Department of Environmental Quality (DEQ), Air Quality Division (AQD), conducted a review of the results from the TO-15 SUMMA canister method Phase I air sampling conducted on November 28th and 29th, 2018 on and around Viant Medical, Inc.'s property located at 520 Watson SW, Grand Rapids, Michigan. These results were received by the DEQ on December 21st, 2018 and January 2nd, 2019. The purpose of this data review was to determine Viant Medical, Inc.'s compliance with the requirements of the federal Clean Air Act; Part 55, Air Pollution Control, of the Natural Resources and Environmental Protection Act, 1994 PA 451, as amended (Act 451); and the Air Pollution Control Rules.

During the data review, staff noted the following:

| Process Description | Rule/Permit Condition Violated | Comments |
|--------------------------------------|--------------------------------|--|
| Ethylene Oxide Sterilization Process | Rule 901(a) | Failure to meet the requirements of Rule 901(a). |

The sampling results identified elevated concentrations of ethylene oxide in the air around Viant Medical, Inc. The Phase I TO-15 SUMMA canister ethylene oxide sampling results are enclosed for your review.

Please initiate actions necessary to correct the cited violation and submit a written response to this Violation Notice by January 18, 2019. The written response should include: the dates the violation occurred; an explanation of the causes and duration of the violation; whether the violation is ongoing; a summary of the actions that have been taken and are proposed to be taken to correct the violation and the dates by which these actions will take place; and what steps are being taken to prevent a reoccurrence.

Please submit the written response to the DEQ, AQD, Grand Rapids District, at 350 Ottawa Avenue NW, Unit 10, Grand Rapids, Michigan 49503 and submit a copy to


Mr. Bryan Curry
Viant Medical, Inc.
Page 2
January 4, 2019

Ms. Jenine Camilleri, Enforcement Unit Supervisor at the DEQ, AQD, P.O. Box 30260,
Lansing, Michigan 48909-7760.

If Viant Medical, Inc. believes the above observations or statements are inaccurate or do not constitute violations of the applicable legal requirements cited, please provide appropriate factual information to explain your position.

Thank you for your attention to resolving the violation cited above. If you have any questions regarding the violation or the actions necessary to bring this facility into compliance, please contact me at the number listed below.

Sincerely,



April Lazzaro
Senior Environmental Quality Analyst
Air Quality Division
616-558-1092

Enclosure

cc: Ms. Mary Ann Dolehanty, DEQ
Dr. Eduardo Olaguer, DEQ
Mr. Christopher Ethridge, DEQ
Ms. Jenine Camilleri, DEQ
Ms. Heidi Hollenbach, DEQ